



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/774,135	02/06/2004	Lukas Eisermann	PC806.00/31132.121	8402
46333	7590	05/14/2010		
Medtronic			EXAMINER	
Attn: Noreen C. Johnson, IP Legal Department			COMSTOCK, DAVID C	
2600 Sofamor Danek Drive				
Memphis, TN 38132			ART UNIT	PAPER NUMBER
			3733	
			MAIL DATE	DELIVERY MODE
			05/14/2010	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/774,135	Applicant(s) EISERMANN ET AL.	
	Examiner DAVID COMSTOCK	Art Unit 3733	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 06 January 2010.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-30 and 37-44 is/are pending in the application.
- 4a) Of the above claim(s) 11 and 12 is/are withdrawn from consideration.
- 5) ☒ Claim(s) 37-40 is/are allowed.
- 6) ☒ Claim(s) 1-10, 13-30 and 41-44 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 06 February 2004 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☒ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|---|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-10, 13-30 and 41-44 are rejected under 35 U.S.C. 103(a) as being unpatentable over Erickson et al. (6,368,350; of record) in view of Ferree (6,706,068).

Erickson et al. disclose an implant comprising first and second components 20 and 21 (see, e.g., fig. 7). The components comprise a convex projection and a concave recess, 25, 46. The projection and recess are offset from the center. The implant comprises a bone-growth promoting coating (see, e.g., col. 6, lines 61-64). The implant can be formed of alloys such as cobalt/chromium, etc. (see, e.g., col. 5, lines 53-55). The lip of the lower component (i.e., nearest reference numeral 26 in Fig. 7) forms a notch that is capable of being grasped by an instrument such as forceps. While spikes 29 are shown in Fig. 7, Erickson et al. teach that the implant can include any of various features for facilitating anchoring of the implant to the vertebrae (see, e.g., col. 5, lines 40-48). Thus, Erickson et al. discloses the claimed invention except for the flanges. Ferree discloses providing flanges longitudinally on bearing surfaces of an articulating implant in order to securely fix the implant to the vertebrae (see, e.g., Fig. 4; col. 3, lines 3-4; and col. 4, line 26). The flanges have holes therein (see Fig. 4). It would have

Art Unit: 3733

been obvious to one having ordinary skill in the art at the time the invention was made to provide the implant of Erickson et al. with longitudinally extending flanges having holes therein, in view of Ferree, in order to securely fix the implant to the vertebrae. At least the linear edges of the flanges are substantially sharp, and the flanges extend parallel to each other. Regarding claims 9, 10 29 and 30, it would have been further obvious to provide an additional duplicate flange on each of the top and bottom portions of the implant, e.g., to more securely fix the implant to the vertebrae, since it has been held that mere duplication of the essential working parts of a device involves only routine skill in the art. *St. Regis Paper Co. v. Bemis Co.*, 193 USPQ 8. With regard to statements of intended use and other functional statements, they do not impose any structural limitations on the claims distinguishable over the applied art, which is capable of being used as claimed if one so desires. *In re Casey*, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 136 USPQ 458, 459 (CCPA 1963). Furthermore, the law of anticipation does not require that the reference “teach” what the subject patent teaches, but rather it is only necessary that the claims under attack “read on” something in the reference. *Kalman v. Kimberly Clark Corp.*, 218 USPQ 781 (CCPA 1983). Furthermore, the manner in which a device is intended to be employed does not differentiate the claimed apparatus from prior art apparatus satisfying the claimed structural limitations. *Ex parte Masham*, 2 USPQ2d 1647 (1987). The implant of Erickson et al. (as modified by Ferree) has an offset configuration as set forth above and there is nothing to prevent, e.g., the upper portion from being turned around in the other direction such that the components are offset in opposite directions, if desired. Thus, the implant components

Art Unit: 3733

are inherently capable of being used in an opposed offset manner and could be used to accomodate a spondylosed relationship between adjacent vertebrae.

Allowable Subject Matter

Claims 37-40 are allowable over the prior art of record.

Conclusion

Applicant's amendment, adding the new limitations to claims 1 and 13 and adding new claims 41-44, necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to David Comstock whose telephone number is (571) 272-4710 (a detailed message should be left if Examiner is unavailable). If attempts to

Art Unit: 3733

reach the Examiner by telephone or voicemail are unsuccessful, the examiner's supervisor, Eduardo Robert, can be reached at (571) 272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

/David Comstock/
Examiner, Art Unit 3733

/Eduardo C. Robert/

Supervisory Patent Examiner, Art Unit 3733